



AQUA PRO-TECH LABORATORIES  
Certified Environmental Testing



# ANALYTICAL RESULTS

STANDARD DELIVERABLES FORMAT

APL WORK ORDER NUMBER: 9060240

Lyons Environmental Services, LLC

Project: Lafayette School

Brian Wood  
Laboratory Director

All Results meet the requirements of the National Environmental Laboratory Accreditation Conference and/or State specific certifications as applicable.



AQUA PRO-TECH LABORATORIES  
 Certified Environmental Testing

## Analytical Results Summary

Lafayette School  
 9060240-02 (Drinking Water)

WC #8

<b>Collected</b> 06/06/2019 06:45	<b>Received</b> 06/06/2019 13:10	<b>Contact</b> Carrie-Anne Lyons
--------------------------------------	-------------------------------------	-------------------------------------

Lab Section/ Analysis	Method	Prepared	Analyzed	Result	Qual	MDL	RL	Units
<b>Total Metals</b>								
Copper	EPA 200.8	06/19/19 22:56	06/19/19 22:56	0.0977			0.00200	mg/L
Lead	EPA 200.8	06/19/19 22:56	06/19/19 22:56	ND	U		0.00200	mg/L

FootNotes

RL - Reporting limit  
 MDL - Minimum detection limit  
 ND - Indicates compound analyzed for but not detected  
 J - Indicates estimated value

B - Indicates compound found in associated blank  
 E - Concentration exceeds highest calibration standard  
 D - Indicates result is based on a dilution  
 P - Greater than 25% diff. between 2 GC columns.  
 H - Indicates a Hold Time violation



AQUA PRO-TECH LABORATORIES  
 Certified Environmental Testing

## Analytical Results Summary

Lafayette School  
 9060240-04 (Drinking Water)

WC #11


Collected 06/06/2019 06:49	Received 06/06/2019 13:10	Contact Carrie-Anne Lyons
-------------------------------	------------------------------	------------------------------

Lab Section/ Analysis	Method	Prepared	Analyzed	Result	Qual	MDL	RL	Units
<b>Total Metals</b>								
Copper	EPA 200.8	06/19/19 23:04	06/19/19 23:04	0.0421			0.00200	mg/L
Lead	EPA 200.8	06/19/19 23:04	06/19/19 23:04	ND	U		0.00200	mg/L

FootNotes

RL - Reporting limit  
 MDL - Minimum detection limit  
 ND - Indicates compound analyzed for but not detected  
 J - Indicates estimated value

B - Indicates compound found in associated blank  
 E - Concentration exceeds highest calibration standard  
 D - Indicates result is based on a dilution  
 P - Greater than 25% diff. between 2 GC columns.  
 H - Indicates a Hold Time violation



www.aquaprotechlabs.com

1275 BLOOMFIELD AVENUE • BUILDING 0  
FAIRFIELD, NEW JERSEY 07004

TEL: 973.227.0422  
FAX: 973.227.2813

CONTAMINATION LEVEL  
 LOW  
 MEDIUM  
 HIGH

CHAIN OF CUSTODY

CLIENT: Lyn3 Env.  
 ADDRESS: \_\_\_\_\_  
 PHONE: 32-566-0038  
 E-MAIL: \_\_\_\_\_  
 SEND REPORT TO: \_\_\_\_\_  
 ADDRESS: \_\_\_\_\_  
 PHONE: \_\_\_\_\_  
 SEND INVOICE TO: \_\_\_\_\_  
 ADDRESS: \_\_\_\_\_  
 PROJECT NAME: Lafayette School  
 PROJECT MGR: \_\_\_\_\_  
 PROJECT or PO #: \_\_\_\_\_  
 SAMPLED BY: GG

TURN-AROUND TIME  
 APL STANDARD 2 weeks  
 RUSH (choose one below, may need lab approval)  
 24 hr. date & time required  
 48 hr. date & time required  
 72 hr. date & time required  
 1 week

REPORT/ELECTRONIC FORMAT  
 RESULTS ONLY  
 NJ DEP REDUCED  
 NJ DEP FULL  
 STATE FORMS/2 REPORTING  
 EMAIL DELIVERY  
 HAZSITE EDD  
 EXCEL  
 SRP#: 1913502

MATRIX ABBREVIATIONS: D - DRINKING WATER G - GROUNDWATER W - WASTEWATER S - SOIL SL - SLUDGE C - CONCRETE L - LAKE

APL Lab ID#	Sample Source: Field ID	Date	Time	Sample Type				Preservative	Analysis Requested
				A	B	C	M		
9060240-01	WC #1	6/6/19	6:40 AM	X					Pb + Cu
-02	WC #8	"	6:45 AM	X					↓
-23	WC #9	"	6:34 AM	X					↓
-24	WC #11	"	6:49 AM	X					↓
-25	WC #13	"	6:52 AM	X					↓

RELINQUISHED BY: Steve Lyons SIGN Steve Lyons DATE & TIME 6/6/19  
 RECEIVED BY: Depez SIGN Depez DATE & TIME 1/3/10  
 RELINQUISHED BY: \_\_\_\_\_ SIGN \_\_\_\_\_ DATE & TIME \_\_\_\_\_  
 RECEIVED BY: \_\_\_\_\_ SIGN \_\_\_\_\_ DATE & TIME \_\_\_\_\_

COMMENTS/SPECIAL INSTRUCTIONS

\_\_\_\_\_

Cooler Temp. upon receipt at lab 4.0